

EVALUATION REPORT (AS PER RULE 35 OF PUBLIC PROCUREMENT RULES, 2017)

1	Name of Procuring Agency	SKBZ Hospital/ CMH Muzaffarabad					
2	Method of Procurement	BY PPRA-AK Open Competitions Bidding					
3	Title of Procurement	Repair of Electro Medical Equipments					
4	Tender Inquiry No.	24 Nov 2018, Daily Jahan Pakistan, Daily Shammal Muzaffarabad AJK-184 N					
5	Date & Time of Bid Closing	Nil					
6	Date & Time of Bid Opening	10-Dec- 2018 at 12:00h					
7	No. Of Bids Received	10- Dec- 2018 at 12:30h					
8	Criteria For Bid Evaluation	Only Six Bidders					
9	Details of Bid (s) Evaluation	Technical/ Financial					
S.N	Name of company	Submitted Technical offer for equipments	Marks		Total evaluated cost 4724800.00	Unit Price (for)	Rules / regulation /SBD/ policy basis for rejection/ acceptance as per rules 35 of AJK PPRA Rules, 2017
			Technical	Financial			
01	M/S Medi Aid Muzaffarabad	Auto clave (Unister)	95%	425000.00	-	425000.00	Accepted being technically responsive & on single basis.
	//	Ultra Sound Probe HS2000 Honda	//	175000.00	-	175000.00	Accepted being technically responsive & on single basis.
	//	Washing Machine Model WB4140	//	325000.00	-	325000.00	Accepted being technically responsive & on single basis.
	//	Washing Machine Model WB 413011	//	345000.00	-	345000.00	Accepted being technically responsive & on single basis.
	//	Anesthesia Machine Ventilator (Frontline Bleese UK/USA)	//	290000.00	-	290000.00	Rejected being highest
02	M/S Four Engineer Rawalpindi	Anesthesia Machine Ventilator (Frontline Bleese UK/USA)	70%	175000.00	-	175000.00	Accepted being technically responsive & on lowest basis.
		Image Intensifier Phillips BV Libra	//	590000.00	-	590000.00	Rejected being highest
		X-Ray Machine (Vella System)	//	115000.00	-	115000.00	Accepted being technically responsive & on lowest basis.

13	M/S Pak Power Moves Faisalabad	Argon Laser (Vital/ Quanted France)	//	258800.00	-	258800.00	Accepted being technically responsive & on single basis.
		Baby Warmer (Satis Medipema France)	//	185000.00	-	185000.00	Accepted being technically responsive & on single basis.
		Incubator	//	390000.00	-	390000.00	Accepted being technically responsive & on single basis.
		Tronequet (ATS 3000 Zimmer)	//	3750000.00	-	3750000.00	Accepted being technically responsive & on single basis.
		X-Ray film Processor (Clasic TOS)	94%	250000.00	-	250000.00	Accepted being technically responsive & on single basis.
//		Gas Plant (Atlas Capco)		300000.00	-	300000.00	Accepted being technically responsive & on single basis.
//		Generator 2500 A Para Saqi	//	230000.00	-	230000.00	Accepted being technically responsive & on single basis.
//		Pulse Oximeter (Sure Sign)	//	70,000.00		70,000.00	Accepted being technically responsive & on single basis.
//		Mobile X-ray machine (Poley Mobile Three)	//	350,000.00		350,000.00	Accepted being technically responsive & on single basis.
04	M/S Human Health Care Lahore	Image Intensifier Phillips BV Libra	95%	571,000.00		571000.00	Accepted being technically responsive & on lowest basis.
		X-Ray Machine (Vella System)	//	578,000.00	-	578000.00	Rejected being highest
		Mobile X-ray machine (Poley Mobile Three)	//	797000.00	-	797000.00	Rejected being highest

MEDICAL OXYGEN GAS

S.N	Name of company	Submitted Technical offer for item	Marks		Total evaluated cost 500000/-	Unit Price (for)	Rules / regulation /SBD/ policy basis for rejection/ acceptance as per rules 35 of AJK PPRA Rules, 2017
			Technical	Financial			
1.	M/S MEDIGAS PRIVATE LIMITED LAHORE	Medical oxygen gas	90&	175.714 per cubic meter		175.714 per cubic meter	Accepted being technically responsive & single bidder

M. S. Ahmed
 (Dr. Mehboob Ahmed Sheikh)
 Additional Medical Superintendent
 SKBZ Hopsital / CMH Muzaffarabd

W

Ultrasonid Machine	Model: HS 2000 Honda Faults: Convex probe faulty (Need new)	Not Quoted	<u>Rs. 175000/-</u>	Not Quoted	Not Quoted	Not Quoted
Washing Machine	Model: W4240H VFD + Wiring + programming need Install	Not Quoted	<u>Rs. 325000/-</u>	Not Quoted	Not Quoted	Not Quoted
Pulse Oximeter	Model: Suresign vs3	Not Quoted	Not Quoted	<u>Rs. 70000/-</u>	Not Quoted	Not Quoted
Mobile X-Ray Machine	Model: (Poly mobile 3)	<u>Rs. 797000/-</u>	Not Quoted	<u>Rs. 350000/-</u>	Not Quoted	Not Quoted
13. Washing Machine	Model: WB413011	Not Quoted	<u>Rs. 345000/-</u>	Not Quoted	Not Quoted	Not Quoted
14. Baby Warmer	Model: Satis medi prema france	Not Quoted	Not Quoted	Not Quoted	Not Quoted	<u>Rs. 185000/-</u>
15. Incubators	Model: Satis medi prema france	Not Quoted	Not Quoted	Not Quoted	Not Quoted	<u>Rs. 390000/-</u>
16. Tourniquate	Model: AST3000 Zimmer	Not Quoted	Not Quoted	Not Quoted	Not Quoted	<u>Rs. 375000/-</u>

Recommendation

- Rate quoted by M/S Human Health Care Lahore, S. No 04 is being accepted on lowest basis.
- Rate quoted by M/S Medi- Aid Muzaffarabad S. No 1,9,10 & 13 are being accepted on single quotation
- Rate quoted by M/S Pak Power Moves Faisal Abad S. No5,7,8,11 & 12 are being accepted on single basis.
- Rate quoted by M/S four Engineer Rawalpindi S. No2,14 & 15 are being accepted on single basis and, 3&6 are being accepted on lowest basis.

HOD/IC Eye Dept _____

IC Gas Plant _____

IC Generator Room _____

HOD/IC Laundry _____

HOD/IC Radiology _____

HOD /IC Ortho Dept _____

HOD/IC OT _____

HOD/IC Gynec _____

(Irum Maqsood) _____
Electro Medical Engineer
SKBZ Hosp/Ak CMH Muzaffarabad

(Dr. Noman Manzoor butt) _____
Deputy Medical Superintendent
SKBZ Hosp/Ak CMH Muzaffarabad

(Col. Zulfiqar Ali Khan) _____
Deputy Comdt SKBZ Hosp
/Ak CMH Muzaffarabad

(Dr. Tanveer Shafi) _____
Director Health Services (Procurement)

APPROVED BY
Secretary Health
ASJK

Brig
Comdt/MS
Hassan Iqbal
SKBZ Hosp/AKCMHMuzaffarabad

COMPARATIVES STATEMENT OF REPAIR ELECTRO MEDICAL EQUIPMENTS SKBZ HOSPITAL AK CMH MUZAFFARABAD FOR F/Y 2019 '19

	NAME OF ITEMS	(1)	(2)	(3)	(4)	(5)
		M/S Humen Health Care Lahore	M/S Medi-Aid Muzaffarabad	M/S Pak Power Moves Faisal Abad	M/S Hospicare Systems Karachi	M/S Four Engineer Rawalpindi
	Auto Clave (OT) Model: <u>Unister</u> Faults: Vaccum Pump Faulty, steam generator faulty & complete over hauling needed	Not Quoted	Rs. 425000/-	Not Quoted	Not Quoted	Not Quoted
2.	Argon laser (Eye) Model: <u>Vital/Quanted france</u> Faults: Touch screen need replacements	Not Quoted	Not Quoted	Not Quoted	Not Quoted	Rs. 350000/-
3.	Anesthesia Machine Ventilator Model: <u>Frontline Blease UK/USA</u> Faults: Accessories O2 Cell Need Replacement/Bellow ,Gauges need replacement, calibration/ touch screen, cct board	Not Quoted	<u>Rs. 290000/-</u>	Not Quoted	Not Quoted	<u>Rs. 175000/-</u>
4.	Image Intensifier Model: <u>Philips BV libra</u> Faults: Problems in imaging/ Display board faulty	<u>Rs. 571000/-</u>	Not Quoted	Not Quoted	Not Quoted	<u>Rs. 590000/-</u>
5.	X Ray Film Processor (Radiology) Model: <u>Classic EOS</u> Faults: Need Complete over hauling	Not Quoted	Not Quoted	Rs. 250000/-	Not Quoted	Not Quoted
6.	X Ray Machine (Radiology) Model: <u>Vella System</u> Faults: Control Console Faulty	<u>Rs. 578000/-</u>	Not Quoted	Not Quoted	Not Quoted	<u>Rs. 115000/-</u>
7.	Gas Plant (Compressor) (Gas Plant) Model: <u>Atlas Capco</u> Faults: <u>(Oil Replacement)</u> <ul style="list-style-type: none"> • <u>Changing</u> • <u>Filter</u> • <u>Oil Splitters Changing</u> • <u>Line filters need replacement</u> • <u>Manifold control system (01)</u> • <u>Regulatory system (Reg) water, Pressure auto and cturming system need repairing with maintenance</u> • <u>Sensor System of dryer fan (02)</u> 	Not Quoted	Not Quoted	Rs. 300000/-	Not Quoted	Not Quoted
8.	Generator Room Breaker: 2500A TeraSaki (Wapda Side) Faults: Need repairing/replacing of mechanical parts.	Not Quoted	Not Quoted	Rs. 230000/-	Not Quoted	Not Quoted

Iram Shahzadi
Anesthesiologist SKBZH/CMH
Asst. Professor AJKMC, M.L

Comparatives statement for Repair EUP1 2018-19 (1)

Condt/mz